

Investigation of Discordant EQA Findings

Stephen Hill

Disclosures of Interest and Permissions

- Employed by Hamilton Health Sciences, Hamilton
- Committee member for Quality Management Program - Laboratory Services
 - QMPLS is the organization responsible for EQA and accreditation for medical laboratories in Ontario.
 - QMPLS material remains the property of QMPLS and may not be used without permission. The reader is cautioned not to take any single item, or part thereof, of this document out of context. Information presented in this document is the sole property and copyright of the Quality Management Program—Laboratory Services (QMP–LS). No part of the material protected by this copyright may be reproduced without written permission from QMP–LS. The logos and/or symbols used are the property of QMP–LS or other third parties. You are not permitted to use these logos and/or symbols without written permission from QMP–LS or such third party that may own the logos and/or symbols. Copyright ©2012 Quality Management Program—Laboratory Services, a department of the Ontario Medical Association.
 - All rights reserved.

Learning Objectives

- Understand the importance of EQA in maintaining the quality of the laboratory service.
- Identify the steps required in an investigation of EQA discordant results.
- Identify and discuss examples of good, adequate and inadequate investigations

- You have just received notice from your EQA provider that you have a “discordant result”. You are now required to perform an investigation and submit the results to your EQA provider
- What do you do? See the next slide and choose all appropriate responses.

1. Curse those guys from the EQA organization as a bunch of nosey busybodies that have no clue what goes on in a busy Clinical Chemistry Laboratory.
2. Ignore it – this evidence is contrary to your own evidence of good laboratory performance.
3. Call your instrument/reagent vendor and ask that a service man be dispatched immediately.
4. Call your instrument/reagent vendor and ask for their advice on how to respond.
5. Undertake a thorough and systematic investigation to identify the root cause, then develop and institute a corrective plan to ensure that this problem is unlikely to occur again.


How to Examine a Problem

- First – clearly understand the problem that needs to be solved
- Second – put in place measures that solve the problem and prevent it from happening again
- Third – continue to monitor the process to be sure that your solution has worked.

“There is a mouse eating the cheese”

Ask why 5 times

Dig deeper in to the problem until the real problem is identified.



5 Whys Root Cause Analysis Worksheet – A Back to the Basics Improvement Template

5 Why's Worksheet

Define the Problem:

Why is it happening?

1.

Why is that?

2.

Why is that?

3.

Why is that?

4.

Why is that?

5.

Caution: If your last answer is something you cannot control, go back up to previous answer.




This document is the intellectual property of The IPL Group LLC and is covered under the international Copyright Convention © 2007. However, it is provided as a free template for organizations to use in the improvement of their internal and external business operations. Go to www.TheIPLGroup.com to download other free templates to improve performance.

Circular Approach

Encourages continuous problem solving

Did putting a latch on the screen door keep the mice out?



ROOT CAUSE ANALYSIS WORK SHEET

Pre-analytical		
Possible Cause	What to Check	Remedial Action
Analytical		
Possible Cause	What to Check	Remedial Action
Post-Analytical		
Possible Cause	What to Check	Remedial Action

QMPLS Approach

- A worksheet is attached to every discordant finding
- Contains a series of questions that are intended to
 - Help the laboratory with the root cause analysis.
 - Reassure QMPLS of the quality processes of the laboratory in question.
- Extra vials of survey material may be available to laboratories.

QMPLS Approach

1. Was the sample re-tested after receipt of the survey results? If yes, please describe findings.
1. Was a replacement sample requested from QMPLS? If yes, please describe findings.
2. Please comment on your investigation of what happened.
3. Please indicate root or contributing cause(s) of problem, if known.
4. Please identify corrective and preventive action(s).
5. Enter Discordant Finding code.

Chemistry
Investigation of Discordant
External Quality Assessment
Finding – Page 1
 Survey:

QMPLS Coding System

Coding the problem allows QMPLS to look for patterns and common problems that may require attention at a higher level+ level.

Most commonly used codes

RANDOM

UNEXPLAINED

Continued unwarranted use of the “unknown” or “random error” codes is sure to raise eyebrows at QMPLS.

“Random error” is an acceptable response if the investigation clearly identifies that the error is random.

“Unknown” is not an acceptable response for a poorly done investigation. If a thorough and systematic investigation that rules out most (if not all) causes of the error, “unknown” is a reasonable response.

Due at QMP–LS by:

Your discordant result(s):

Analyte / Test	Vial(s)	Your Result(s)	Assigned Value(s)	Acceptable Range(s)	PAD / Q Score	Action

Please answer the following questions:

1. Was the problem related to any of the following? Select as many as are applicable to this situation. (Please explain on page 2).

Code	A. Clerical activities associated with test	Code	C. Method(s) used to perform the test...(Cont'd)
C01	<input type="checkbox"/> Computer keying/Data entry error	M05	<input type="checkbox"/> Inadequate or inappropriate QC program
C02	<input type="checkbox"/> EQA Analysis Worksheet difficult to use	M06	<input type="checkbox"/> Inadequate written procedure/Procedure poorly presented
C03	<input type="checkbox"/> Failure to inform QMP-LS of method change, therefore analysed in inappropriate peer group	M07	<input type="checkbox"/> Inadequate LIS reporting function
C06	<input type="checkbox"/> Incorrect analyte reported	M09	<input type="checkbox"/> Inappropriate QC material
C12	<input type="checkbox"/> Incorrect units reported or misplaced decimal point	M12	<input type="checkbox"/> Manufacturer or supplier related issue e.g., problem with reagent performance
C14	<input type="checkbox"/> Mislabeled vial, slide or test tube	M13	<input type="checkbox"/> Method affected by temperature in laboratory
C15	<input type="checkbox"/> Result not correctly transcribed to Analysis Worksheet	M14	<input type="checkbox"/> Method inaccurate/imprecise
C16	<input type="checkbox"/> Result omitted on Analysis Worksheet	M15	<input type="checkbox"/> Method lacks sensitivity/specificity
C17	<input type="checkbox"/> Result reported for the wrong sample	M16	<input type="checkbox"/> Method not validated in-house
C99	<input type="checkbox"/> Other clerical problem (please specify on page 2)	M17	<input type="checkbox"/> Method subjected to interference
Code	B. Technical operation of method	M20	<input type="checkbox"/> Problem with water supply
T02	<input type="checkbox"/> Calculation and/or conversion error	M22	<input type="checkbox"/> Quality control material not at relevant concentrations
T03	<input type="checkbox"/> Dilution error	M23	<input type="checkbox"/> Reference interval not validated for relevant clinical levels
T04	<input type="checkbox"/> EQA material improperly prepared or stored	M25	<input type="checkbox"/> Scheduled instrument maintenance not performed appropriately
T05	<input type="checkbox"/> Expired kits or reagents used	M99	<input type="checkbox"/> Other method problem (please specify on page 2)
T06	<input type="checkbox"/> Failure to act on inappropriate QC results	Code	D. Equipment Function
T07	<input type="checkbox"/> Failure to add reagent or sample to test system	E02	<input type="checkbox"/> Hardware problem
T08	<input type="checkbox"/> Failure to calibrate after major instrument failure	E03	<input type="checkbox"/> Detection system error
T09	<input type="checkbox"/> Failure to correctly interpret data provided on Analysis Worksheet	E04	<input type="checkbox"/> Electrical interference
T10	<input type="checkbox"/> Failure to follow instructions on Analysis Worksheet	E05	<input type="checkbox"/> Instrument software error/problem
T11	<input type="checkbox"/> Failure to follow written procedures	E06	<input type="checkbox"/> Insufficient aspiration of sample
T15	<input type="checkbox"/> Failure to treat EQA sample in a routine manner	E07	<input type="checkbox"/> Obstruction of instrument's tubing or aperture
T16	<input type="checkbox"/> Inability of staff to apply knowledge to situation	E08	<input type="checkbox"/> Problem with equipment function/Analyzer defect
T18	<input type="checkbox"/> Inadequate equipment maintenance	E09	<input type="checkbox"/> Problem with instrument data processing functions
T19	<input type="checkbox"/> Inadequate mixing of sample	E99	<input type="checkbox"/> Other equipment problem (please specify on page 2)
T20	<input type="checkbox"/> Inappropriate handling of samples	Code	E. Organizational Factors
T23	<input type="checkbox"/> Incorrect instrument calibration	O01	<input type="checkbox"/> Inadequate and/or inappropriate equipment
T25	<input type="checkbox"/> Instrument error message misinterpreted/overlooked	O02	<input type="checkbox"/> Inadequate communication with all rotating staff
T26	<input type="checkbox"/> Manufacturer's instructions not followed	O04	<input type="checkbox"/> Inadequate educational in-service provided for staff
T32	<input type="checkbox"/> Pipette not appropriately calibrated	O05	<input type="checkbox"/> Inadequate workplace design
T34	<input type="checkbox"/> Samples mixed up on bench or wrong sample tested	O06	<input type="checkbox"/> Insufficient staffing
T36	<input type="checkbox"/> Test result misinterpreted	O07	<input type="checkbox"/> Insufficient training to perform task
T37	<input type="checkbox"/> Testing delayed after reconstitution	O09	<input type="checkbox"/> Lack of organizational awareness or prioritization
T38	<input type="checkbox"/> Testing delayed after aliquoting sample on to instrument	O10	<input type="checkbox"/> Staff not qualified to perform the task
T99	<input type="checkbox"/> Other technical problem (please specify on page 2)	O11	<input type="checkbox"/> Supervision not available
Code	C. Method(s) used to perform the test	O99	<input type="checkbox"/> Other organizational factors (please specify on page 2)
M02	<input type="checkbox"/> Carryover from previous sample	Code	F. Unexplained
M03	<input type="checkbox"/> Erroneous manufacturer calibrator value	U01	<input type="checkbox"/> Random
		U02	<input type="checkbox"/> Unexplained
		U99	<input type="checkbox"/> Other explanation (please specify on page 2)

QMPLS Scoring system is based on Percent Allowable Difference, or PAD

$$\text{PAD} = ([x - x_a]/\text{APL}) \times 100$$

Where x is the participant response, x_a is the “Assigned Value,” and APL is the Allowable Performance Limit.

	Current Survey Rules	Action Required
	$ \text{PAD} \leq 100$	No Action Required
W1	$ \text{PAD} > 100 \ \& \ \leq 150$ for single result	Investigate and retain documents internally
A1	$ \text{PAD} > 150$ for single result	Investigate and submit to QMPLS
A2	$ \text{PAD} > 100$ for ≥ 2 results on same analyte	Investigate and submit to QMPLS
A3	$ \text{PAD} > 100$ for ≥ 2 results on same analyte in last 2 of 3 surveys	Investigate and submit to QMPLS

Example # 1


Manufacturer Group: Ortho Clinical Diagnostics

Analyte: Total Calcium

Vial #	Analyte	Result	Target	Range	PAD Score	Action Code
1	Calcium	2.40 mmol/L	2.23	2.12 – 2.34	155	A2
2	Calcium	3.36	3.05	2.90 – 3.20	207	A2
3	Calcium	2.40	2.26	2.15 – 2.37	127	A2


Calcium, Total - VIAL 1

N: 177 Grouped By Reagent Assigned Value: 2.23 Allowable Limits: 2.12 - 2.34



Calcium, Total - VIAL 2

N: 177 Grouped By Reagent Assigned Value: 3.05 Allowable Limits: 2.90 - 3.20



Calcium, Total - VIAL 3

N: 177 Grouped By Reagent Assigned Value: 2.26 Allowable Limits: 2.15 - 2.37



INVESTIGATION SUMMARY

Investigation

- Review of calibration and QC data determined results were acceptable.
- Re-calibration using different Calibration lot # but same slide lot #.
- Re-assay of the QMPLS vials – no change in values.
- After consultation with manufacturer, major maintenance performed.
- Before and after patient correlation revealed a slight drop in results. QMPLS vials retested; results now acceptable.
- Manufacturer has decreased routine maintenance visits from every six months to once per year. Several key maintenance procedures that must be done every six months are now being done only every year.

Sample Retest: After major maintenance and re-calibration

Vial 1	2.26 mmol/L	All results now in acceptable range
Vial 2	3.14 mmol/L	
Vial 3	2.28 mmol/L	

Root Cause: Change in maintenance schedule resulted in failure to perform critical procedures.

Corrective action: Items performed by service rep yearly will be performed by laboratory staff every 6 months.

DF Code: T18 Inadequate equipment maintenance

Example #2

Calcium


Vial #	Analyte	Result	Target	Range	PAD Score	Action Code
1	Calcium	2.36 mmol/L	2.23	2.12 – 2.34	118	A2
2	Calcium	3.27	3.05	2.90 – 3.20	147	A2

Manufacturer Group – Ortho Clinical Diagnostics

Note: Calcium target values are determined by analysis at 2 external reference laboratories.


Calcium, Total - VIAL 1

N: 177 Grouped By Reagent Assigned Value: 2.23 Allowable Limits: 2.12 - 2.34




Calcium, Total - VIAL 2

N: 177 Grouped By Reagent Assigned Value: 3.05 Allowable Limits: 2.90 - 3.20



Calcium, Total - VIAL 3

N: 177 Grouped By Reagent Assigned Value: 2.26 Allowable Limits: 2.15 - 2.37



INVESTIGATION SUMMARY

Sample Retest: Vial 1 - 2.33, Vial 2 - 3.19

Investigation:

- Retest results within acceptable “your” acceptable range
- QC was running within acceptable parameters
- No contributing error determined from our investigation

Root Cause: No specific cause noted. Our method running slightly higher than other methods.

Corrective Action: We will continue to monitor QC closely

DF Code: U01 Random

Your comments on this investigation?

Example #3


BLOOD GASES

Multiple instruments tested. Discordant findings noted for 3 instruments. Three separate investigations undertaken and reported.

Instrument #	Vial #	Analyte	Result	Target	Range	PAD Score	Action Code
1	2	pCO ₂	66 mm Hg	79	72 - 86	-186	A3
2	2	pCO ₂	88	79	72 - 86	129	A3
3	1	pCO ₂	52	47	43 - 51	125	A3
3	2	pCO ₂	90	79	72 - 86	157	A3


Manufacturer Group: Siemens Rapidpoint

Vial 1



Siemens Rapidpoint

Vial 2



Siemens Rapidpoint

← Instrument #1

↘ Instrument #2

← Instrument #3

Investigation Summary

	Instrument 1	Instrument 2	Instrument 3
Sample Retest	Retested immediately as per critical value policy 68 mm Hg <div style="border: 1px solid red; padding: 2px; display: inline-block;"><i>Still too low</i></div>	Retested immediately as per critical value policy 80 mm Hg – now within acceptable range	Retested immediately as per critical value policy Vial 1 – 48 mm Hg Vial 2 - 84 mm Hg Both results now within acceptable range
Investigation	Testing performed by POCT operator QC and split sample testing did not show problems Latest instrument calibration verification acceptable	Testing performed by POCT operator QC and split sample testing did not show problems Latest instrument calibration verification acceptable	Testing performed by POCT operator QC and split sample testing did not show problems Latest instrument calibration verification acceptable CAP Proficiency testing acceptable.
Root Cause	Due to high pCO ₂ loss of CO ₂ to room air due to poor specimen handling may have contributed	Unknown – likely due to random error due to high pCO ₂ concentration	Unknown – likely some inaccuracy and imprecision at high pCO ₂ concentrations
Corrective action	Continue to monitor QC & split sample performance Continue to work in conjunction with POCT operators	Continue to monitor QC & split sample performance Continue to work in conjunction with POCT operators	Continue to monitor QC & split sample performance Continue to work in conjunction with POCT operators
DF Code	T99 – other technical error	U 01 – random error	M14 – method inaccurate or imprecise

Example # 4

Cholesterol


Vial #	Analyte	Result	Target	Range	PAD Score	Action Code
1	Cholesterol	7.63	9.42	8.95 – 9.89	-381	A3
2	Cholesterol	6.53	5.96	5.66 – 6.26	190	A3

Manufacturer Group: Ortho Clinical Diagnostics

A3 code indicates flags in 2 of 3 previous surveys


Cholesterol - VIAL 1

N: 129 Grouped By Reagent Assigned Value: 9.42 Allowable Limits: 8.95 - 9.89



Cholesterol - VIAL 2

N: 129 Grouped By Reagent Assigned Value: 5.96 Allowable Limits: 5.66 - 6.26



Note: samples for this survey are single donor, fresh, never frozen samples. Target values established by 2 external reference laboratories

Sample Retest: Repeated using same lot # reagent

Vial #1 – 8.34 mmol/L Still low

Vila #2 – 6.89 mmol/L Still high

Investigation:

Review of histogram shows poor performance by all OCD users.

OCD hotline contacted. Hotline requested fax of laboratory results for this survey

No trouble shooting done with hotline advisor at this time. They informed the lab that OCD specialists were preparing a letter of explanation.

Sent 10 samples to outside laboratory that also uses OCD instrument. Results showed a positive bias across the range of results

Root Cause:

Currently awaiting a letter from OCD that we hope has addressed recent poor performance of Cholesterol on QMPLS LIPS 1101survey.

When received, letter will be reviewed closely and suggestions promptly acted upon. Letter will be uploaded to QMPLS.

Corrective Action:

Two lot #s re-calibrated (current and one other) using same lot # of calibrator.

The 10 samples sent out were reanalyzed using both reagent lot #s
Results correlated well lot-to-lot, better correlation with outside lab but slight positive bias remains.

Will continue to work with OCD and hope that a solution is found.

Patient comparisons between current and new lot #s of reagents will be instituted.

Regular comparison of patient results with peer laboratories will be considered.

DF Code: M12 – manufacturer or supplier related issue e.g. problem with reagent performance

Conclusions

- EQA and external proficiency testing are essential components of a good quality system
- A well performed problem analysis can help to identify issues that affect the quality of your laboratory.
- Corrective action and surveillance can ensure that the problem does not repeat itself.

Thank-you

Further questions or comments?